



399 Reservation Road
Marina, CA 93933 USA
Phone: 831-384-4300
Fax: 831-384-4312
sales@foxthermal.com

Zero CAL-CHECK[®] CERTIFICATE CALIBRATION VALIDATION

Zero CAL-CHECK[®] Performed on: May 03 2019

9:35:22

Firmware version: FT1 V5.5

Fox Meter Serial Number: F04001

Zero CAL-CHECK[®] Results: PASS

Zero CAL-CHECK[®]: 0.46%

Test Temperature 75.3 F

Tag #/Meter Location: *Compressor #528*

Test performed by: *John Doe*

Additional Comments: _____

Zero CAL-CHECK[®] is a calibration routine that validates the flow meter's calibration accuracy by testing the following:

- * Sensor free of residue or film
- * Zero Stability of the meter
- * Thermal Conductivity (heat transfer) repeatability of the sensor

At the conclusion of the test, the meter will display a pass/fail message and the Zero CAL-CHECK[®] data.

A "pass" result confirms the meter is measuring accurately.