



CE Declaration of Conformity

No. F-158 Rev. C  
Fox Thermal Instruments  
399 Reservation Road  
Marina CA, 93933

Fox Thermal Instruments declares that the equipment described below complies fully with the requirements of CE.

Description: Thermal Mass Flow Meter- Model FT1

This equipment has been tested against the following harmonized standards:

2014/30/EU	Emissions and Immunity Testing: EN61326-1:2013
94/9/EC & 2014/34/EU	ATEX Directive; Product Safety Testing: EN 60079-0:2012 + A11:2013, EN 60079-1:2014, EN 60079-31:2014 IECEX scheme: IEC 60079-0:2011, IEC 60079-1:2014, IEC 60079-31:2013

I hereby declare that the Fox Thermal Instruments FT1 has been designed to comply with the relevant sections of the above referenced specifications. The FT1 complies with all applicable Essential Requirements of the Directive.

Signed:

Name: Brad Lesko Title: President/ CEO  
Place: Marina, California Date: 11/27/17