



## EC Declaration of Conformity

No. F-151

Fox Thermal Instruments

399 Reservation Road

Marina CA, 93933

We hereby declare that the equipment described below complies fully with the requirements of the Pressure Equipment Directive 97/23/EC and 2014/68/EU Article 13.

Description: Thermal Mass Flow Meter – Model FT3

Conformity Assessment Modules followed – D1

Name and address of the Notified Body monitoring the manufacturer's quality system: -

BSI, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes, MK5 8PP

Notified Body number - 0086

Harmonized Standards applied: - EN ISO 15614-1 and EN ISO 9606-1

Other technical standards used: - ASME B31.3

Other Community Directives applied: -

2014/30/EU EMC Directive; Emissions and Immunity Testing: EN61326-1:2013

2014/35/EU Low Voltage Directive (LVD); Product Safety Testing: EN61010-1:2010

94/9/EC & 2014/34/EU ATEX Directive; Product Safety Testing: EN 60079-0:2012 + A11:2013, EN 60079-1:2014, EN 60079-31:2014, EN 60529:1991 + A1:2000

IECEX scheme: IEC 60079-0:2011, IEC 60079-1:2014, IEC 60079-31:2013, IEC 60529:2001

I hereby declare that the Fox Thermal Instruments FT3 has been designed to comply with the relevant sections of the above referenced specifications. The FT3 complies with all applicable Essential Requirements of the Directives.

Signed:

Name: Brad Lesko

Title: President/CEO

Place Marina, California

Date 4/8/16