

April 22, 2014

**Fox Thermal Instruments Statement re: NACE MR0175**

The wetted materials of Fox Thermal Instruments' flow meters are made of austenitic stainless steel\*. Austenitic steel is acceptable for use in sour\*\* environments which require NACE MR0175 compliant products.

NACE compliance is dependent upon end user applications. Fox is not responsible for misuse of the meter.

For more information on NACE MR0175, please visit <http://www.bssa.org.uk/topics.php?article=98> .

Sincerely,



Rich Cada  
VP Sales & Marketing  
Fox Thermal Instruments, Inc.

*\* "Austenitic stainless steel" contains the following elements in the specified ranges:*

*C 0.08%max*

*Cr 16%min*

*Ni 8%min*

*P 0.045% max*

*S 0.04%max*

*Mn 2.0%max*

*Si 2.0% max*

*\*\*\*"Sour environments" refers to a gas stream that contains Hydrogen Sulfide (H2S). When mixed with water it becomes Hydrosulfuric Acid.*